



# Agilent Technologies 5348A 26.5 GHz Microwave Counter/ Power Meter

Data Sheet

## Input 1

Frequency Range: 500 MHz to 26.5 GHz

### Sensitivity

500 MHz - 12.4 GHz: -32 dBm (-35 dBm typical)

12.4 GHz - 20.0 GHz: -27 dBm (-32 dBm typical)

20.0 GHz - 26.5 GHz: -20 dBm (-27 dBm typical)

Maximum Input: +7 dBm

Damage Level: +25 dBm, peak

### SWR

500 MHz - 12.4 GHz: <2:1 typical

12.4 GHz - 20.0 GHz: <3:1 typical

20.0 GHz - 26.5 GHz: <3:1 typical

Accuracy:  $\pm 1$  LSD  $\pm$ time base error x frequency

Resolution: 1 Hz or 10 kHz, selectable

## Optional Increased Damage Level Option 006

Protects Input 1 from damage by limiting high level signals.

All specifications are the same except Input 1.

### Damage Level

500 MHz to 6 GHz: +39 dBm (8 watts)

6 GHz to 18 GHz: +36 dBm (4 watts)

18 GHz to 26.5 GHz: +34.8 dBm (3 watts)

### Sensitivity reduced by

3 dBm, 500 MHz to 12.4 GHz

4 dBm, 12.4 GHz to 20.0 GHz

5 dBm, 20.0 GHz to 26.5 GHz

## Input 2

Frequency Range: 10 Hz - 525 MHz

Sensitivity: 25 mV rms (15 mV rms typical)

Impedance: 1 Mohm nominal shunted by <70 pF (10 Hz to 80 MHz) or 50 Ω nominal (10 MHz to 525 MHz)

Damage Level: 50 V or 1 MV

dc - 5 kHz: 250V (dc + ac peak)

>5 kHz: 5.5V rms (+28 dBm) +  $1.25 \times 10^6$  V rms/FREQ

Accuracy:  $\pm 1$  LSD  $\pm [(1.4 \times \text{Trigger Error/Gate Time}) \pm \text{Time Base Error}] \times \text{frequency}$

Resolution: 1 Hz or 10 kHz, selectable

### Automatic Amplitude Discrimination

Automatically measures the largest of all signals present, provided that signal is >6 dB (typical) above any signal within 500 MHz; >20 dB (typical) above any signal within 500 MHz to 26.5 GHz.

### Tracking Speed

Resolution = 1 Hz, Speed = 1 MHz/sec

Resolution = 10 kHz, Speed = 1 GHz/sec

### Acquisition Time

Resolution = 1 Hz, Time = <125 ms

Resolution = 10 kHz, Time = <60 ms

### Maximum Deviation

20 MHz p-p, Automatic mode

60 MHz p-p, Manual mode (via HP-IB only)

Maximum FM Rate: 10 MHz

AM Tolerance: Any modulation index provided the minimum signal level is not less than the sensitivity specification.

### Power Meter Specifications

Frequency Range: 0 MHz to 26.5 GHz, sensor dependent

Power Range: -70 to +20 dBm (100 pW to 100 mW), sensor dependent

Dynamic Range: 50 dB in 10 dB steps

Resolution: 0.01 dB in logarithmic mode, 0.1% of full scale in linear mode.

Auto Filter: The meter automatically selects the required number of averages for the selected range.

### Accuracy

Instrumentation:  $\pm 0.02$  dB or  $\pm 5\%$

Zero Set (digital settability of zero):  $\pm 5\%$  of full scale on most sensitive range. Decrease percentage by a factor of 10 for each higher range,  $\pm 1$  display count.

