



N2795A/N2796A

Single-ended Active Probes

User's Guide



© Copyright Agilent Technologies 2010
All Rights Reserved.



Agilent Technologies

www.valuetronics.com

Contents

Inspecting the Probe ...	3
Cleaning the Probe ...	3
Handling the Probe ...	3
N2795A/N2796A Single-ended Active Probes ...	4
Accessories Supplied with Probe ...	5
Replacement Accessory Kits ...	6
Characteristics and Specifications ...	7
Performance Data Plots ...	9
Safety Information ...	11
Using the N2795A/N2796A Probes ...	12
Dynamic Range and Offset Voltage Limitations ...	14
Dimensions (in millimeters) ...	8
Accessories ...	15
Configurations 1 and 2: Rigid Probe Tips / Spring Probe Tips and Offset Ground / Ground Blade ...	16
Configuration 3: Copper Pads with Flex Ground ...	19
Configuration 4: Y-Lead Adapter with 0.1” Headers ...	20
Configuration 5: Flex Nose Clips ...	21
Configurations 6 and 7: Ground Leads and Right Angle Ground Leads ...	22
Color Coding Rings ...	23
Calibration Testing Procedures ...	24
To Return the Probe to Agilent Technologies for Service ...	36

Inspecting the Probe

- o Inspect the shipping container for damage.

Keep the damaged shipping container or cushioning material until the contents of the shipment have been checked for completeness and the probe has been checked mechanically and electrically.

- o Check the accessories.

- If the contents are incomplete or damaged, notify your Agilent Technologies Sales Office.

- o Inspect the instrument.

- If there is mechanical damage or defect, or if the probe does not operate properly or pass calibration tests, notify your Agilent Technologies Sales Office.
- If the shipping container is damaged, or the cushioning materials show signs of stress, notify the carrier as well as your Agilent Technologies Sales Office. Keep the shipping materials for the carrier's inspection. The Agilent Technologies office will arrange for repair or replacement at Agilent Technologies' option without waiting for claim settlement.

Cleaning the Probe

Disconnect the probe and clean it with a soft cloth dampened with a mild soap and water solution. Make sure the probe is completely dry before reconnecting it to an oscilloscope.

Handling the Probe




Handle the probe with care to avoid injury, especially when it is fitted with the sharp spring or rigid tip. Note that the probe cable is a sensitive part of the probe and, therefore, you should be careful not to damage it through excessive bending or pulling. You should also avoid any mechanical shocks to this product in order to guarantee accurate performance and protection.

N2795A/N2796A Single-ended Active Probes

The N2795A/96A are a new generation of 1 and 2 GHz single-ended active probes with the AutoProbe interface (compatible with Agilent's InfiniVision and Infiniium family of oscilloscopes). These probes integrate many of the characteristics needed for today's general-purpose, high-speed probing - especially in digital system design, component design/characterization, and educational research applications. Its 1M Ω input resistance and extremely low input capacitance (1 pF) provide ultra low loading of the DUT. This, accompanied with superior signal fidelity, makes these probes useful for most of today's digital logic voltages. And with their wide dynamic range (± 8 V) and offset range (± 12 V for N2796A, ± 8 V for N2795A), these probes can be used in a wide variety of applications.

For high signal integrity probing, the N2795A 1 GHz and N2796A 2 GHz active probes are perfect complements to Agilent's 500MHz – 600MHz and 1 GHz bandwidth scopes, respectively. The N2796A 2 GHz probe can also be used with Agilent's 2.5 GHz or higher bandwidth Infiniium scope as a low cost alternative to the InfiniiMax probes.


The N2795A/96A are equipped with a pleasant white LED headlight to illuminate the circuit under test (picture shown below). The probes are powered directly by the InfiniiVision and Infiniium Autoprobe interface, eliminating the need for an additional power supply. The probes also come with a number of accessories that allow for easy connections to the circuit under test (Accessories Supplied with Probe ... page 5)



Accessories Supplied with Probe

The following table shows the standard accessories supplied with the N2795A and N2796A probes. Some of these accessories are discussed in greater detail starting on page 15. The Reference Number for each accessory can be matched to the picture below the table.


Reference Number	Accessory	Quantity
1	Spring probe tip	2
2	Rigid probe tip	2
3	Flex nose clip adapter	1 (red) 1 (black)
4	Copper pads (10x)	1
5	Y-lead adapter, 10 cm	1
6	Right angle ground, 5 cm	1
7	Right angle ground, 10 cm	1
8	Ground blade	2
9	Offset ground	1
10	Flex ground	1
11	Color coding rings (yellow, green, blue, purple)	4 of each color



Replacement Accessory Kits

The following table shows the replacement accessory kits available for the N2795A and N2796A probes. Some of these accessories are discussed in greater detail starting on page 15.

Part Number	Accessory	Quantity
5061-7388	Spring probe tip	2
5061-7389	Rigid probe tip	2
5061-7390 (red)	Flex nose clip adapter	1 (red)
5061-7391 (black)		1 (black)
0960-2908	Copper pads (10x)	1
5061-7392	Y-lead adapter, 10 cm	1
5061-7399	Right angle ground, 5 cm	2
5061-7400	Right angle ground, 10 cm	2
5061-7393	Ground blade	2
5061-7394	Offset ground	2
5061-7397	Flex ground	2
5061-7395	Ground lead, 6 cm	2
5061-7396	Ground lead, 12 cm	2



Characteristics and Specifications

Characteristics and specifications for the N2795A/N2796A Single-ended Active Probes are shown below. The probe / oscilloscope should be warmed up for at least 20 minutes before any testing and the environmental conditions should not exceed the probe's specified limits.

Electrical Characteristics

Attenuation ratio (@ DC)	N2795A: 10:1 +/- 0.5% N2796A: 10:1 +/- 0.5%
Bandwidth (-3 dB)*	N2795A: 1 GHz N2796A: 2 GHz
Probe Risetime (calculated, 10%-90%)	N2795A: 350 ps N2796A: 175 ps
Input Dynamic Range	N2795A: -8V to +8V (DC or Peak AC) N2796A: -8V to +8V (DC or Peak AC)
Input Resistance*	N2795A: 1 M Ω +0%, -2.5% N2796A: 1 M Ω +0%, -2.5%
Non-destructive Max Input Voltage	N2795A: -20V to +20V N2796A: -20V to +20V
Offset Range	N2795A: +/- 8V N2796A: +/- 12 V
DC Offset Error (Output Zero)	N2795A: < +/- 1 mV N2796A: < +/- 1 mV
Flatness (both models)	Typical 0.4 dB (100 kHz - 100 MHz) Typical 0.6 dB (100 MHz - 500 MHz) Typical 0.8 dB (500 MHz - 1 GHz) Typical 1.3 dB (1 GHz - 2 GHz)**
Input Capacitance	N2795A: 1 pF N2796A: 1 pF
Probe Noise (referred to input)	N2795A: < 1 mVrms N2796A: < 1 mVrms
Output Impedance	N2795A: 50 Ω N2796A: 50 Ω
Internal Power	Agilent AutoProbe Interface from Scope (InfiniiVision and Infiniium models)

*Denotes warranted specification after 20 minute warm-up, all others typical

**N2796A only

Dimensions (in millimeters)



Mechanical Characteristics

Approximate Weight (probe only)	100 g
Cable Length	1.3 m

Environmental Specifications

Temperature	Operating: 0 °C to +50 °C Non-operating: -40 °C to +70 °C
Altitude	Operating: 4,000 m (13,000 ft)
Humidity	Operating: 95% room humidity @ 40 °C Non-operating: 90% room humidity @ 65 °C
ESD	8kV HBM

Dimensions (in millimeters)



Performance Data Plots




Figure 1 Frequency response of N2796A (V_{out}/V_{in})

NOTE: This plot is for the spring/rigid probe tip and an offset ground or ground blade.




Figure 2 Input impedance over frequency (red = measured, blue = model)

Safety Information

Warning



To avoid personal injury and to prevent fire or damage to this product or products connected to it, review and comply with the following safety precautions. Be aware that if you use this probe assembly in a manner not specified, the protection this product provides may be impaired.

Handle Probe Tips / Accessories Carefully

Some of the probe tips / accessories are very sharp (the spring tips, for example). You should handle these with care to avoid personal injury.

Use Only Grounded Instruments

Do not connect the probe's ground lead to a potential other than earth ground. Always make sure the probe and the oscilloscope are grounded properly.

Connect and Disconnect Properly

Connect the probe to the oscilloscope and connect the ground lead to earth ground before connecting the probe to the circuit under test. Disconnect the probe input and the probe ground lead from the circuit under test before disconnecting the probe from the oscilloscope.

Observe Probe Ratings

Do not apply any electrical potential to the probe input which exceeds the maximum rating of the probe. Make sure to comply with the voltage versus frequency derating curve found in this manual.

Keep Away From Live Circuits

Avoid open circuitry. Do not touch connections or components when power is present.

Indoor Use Only

Do not operate in wet/damp environments. Keep product surfaces dry and clean.

Do Not Operate With Suspected Failures

Refer to qualified service personnel.

Using the N2795A/N2796A Probes


Using the Headlight

These probes are equipped with two white LED headlights that can be used to illuminate the area around the probe tip. This headlight has a fixed intensity and is turned on/off by pressing the button on the body of the probe (see picture below).



Connecting Accessories

The following picture show where you should connect probe tips and grounding accessories to the body of the probe.



Oscilloscope Compatibility

The N2795A/N2796A probes are compatible with the following Agilent oscilloscopes (those with the Agilent AutoProbe Interface):

- InfiniiVision 5000 Series, software version 6.10 or higher
- InfiniiVision 6000 Series (except 100 MHz model), software version 6.10 or higher
- InfiniiVision 7000 Series, software version 6.10 or higher
- Infiniium 9000 Series, software version 2.50 or higher
- Infiniium 90000 Series, software version 2.50 or higher
- Infiniium 90000 X-Series (requires N5442A adapter)


General Measurement Procedure / Functional Test

The following is a generic measurement procedure you can use to ensure your unit is functioning properly.

- 1 Connect the probe to an oscilloscope channel input and ensure the input impedance of the oscilloscope matches the output impedance of the probe (50Ω).
- 2 Equip the probe with a grounding accessory and connect the ground to the ground terminal on the oscilloscope.
- 3 Connect the probe tip to the Probe Compensation Output on the oscilloscope.
- 4 Enable autoscale on the oscilloscope.
- 5 A square wave should now be displayed on the oscilloscope if the probe is functioning properly.

Dynamic Range and Offset Voltage Limitations

With the N2795A/96A active probes offering 16 Vpp (+/-8V) of dynamic range and +/-12V (N2796A) or +/-8V (N2795A) of offset range, the probes have a wide measuring range of +/-20V, allowing them to be used for a wide variety of applications. For example, if you wanted to measure small AC noise riding on top of a +10V DC signal, you could use the DC offset feature to position the signal at the center of the screen to keep the probe within its dynamic range. You could then zoom in to see the detail at a higher resolution. Trying to measure a signal out of the dynamic range may result in a clipped or distorted waveform.



Accessories

The following sections describe how to use the N2795A/96A probe accessories. They are placed in order from the most recommended configuration to least recommended (based on bandwidth). The configuration card that ships with the probe also shows these recommended configurations.

Configurations 1 and 2: Rigid Probe Tips / Spring Probe Tips and Offset Ground / Ground Blade

Bandwidth with N2795A: 1 GHz

Bandwidth with N2796A: 2 GHz

Replacement Kit Part Numbers (the quantities are the amounts shipped in the replacement kits, not the quantity included with the probe):

Rigid Probe Tip: 5061-7389 (qty. 2)

Spring Probe Tip: 5061-7388 (qty. 2)


Ground Blade: 5061-7393 (qty. 2)

Offset Ground: 5061-7394 (qty. 2)

These rigid and spring-loaded tips are replaceable. To change the probe tip, pull it straight out of its contact socket along the axis of the probe. Once the probe tip is removed, the new tip can be inserted into the contact socket along the axis of the probe. In order to insert the probe tip completely into the housing, carefully press the probe tip against a hard surface.


It is recommended that the spring tips be used with the ground blade as both the tips and ground blade are spring-loaded and will maintain contact with the device under test even when your hand moves slightly.

The Rigid/Spring tips and grounds can be mixed with no effect on performance. It is simply a matter of personal preference. Spring loaded probe tips offer a method of probing signals that is less susceptible to vibration or movement than traditional rigid tips. Many users find it easier to use this type of tip. The spring loaded tips work when they are either partially or fully compressed and are protected against over compression damage.



Both the spring tip and ground blade are spring-loaded

The recommended grounding accessory for the rigid probe tip is the offset ground as shown.




Both of these configurations are the highest performing for these probes. For more information on recommended configurations, consult the configuration card that shipped with your probe.

Warning




You should exercise caution when using these sharp probe tips to avoid personal injury.


The following images show you how to insert the probe tips and grounds.



Dynamic Range and Offset Voltage Limitations



Configuration 3: Copper Pads with Flex Ground



Bandwidth with N2795A: 900 MHz*

Bandwidth with N2796A: 1.8 GHz*

*Highly dependent on the number of ground connections between the copper pad and the device under test

Replacement Kit Part Numbers (the quantities are the amounts shipped in the replacement kits, not the quantity included with the probe):

Copper Pads: 0960-2908 (qty. 10)

Flex Ground: 5061-7397 (qty. 2)

These self adhesive copper pads can be attached on top of an IC and connected to its ground pins to create a convenient ground plane for the probe to connect to. When used with the flex ground, this method provides an ideal ground connection for probing signals with high frequency content. However, to maximize the performance of this setup, you need to solder as many jumper wires as is practical from the ground pins of the package to the copper pad. The more ground connections, the less inductance in the ground path and the higher the bandwidth capability of the measurement. Insert the Flex Ground and bend it into a shape that makes it easy to contact the copper ground pad on top of the package.



Configuration 4: Y-Lead Adapter with 0.1” Headers



Bandwidth with N2795A: 750 MHz



Bandwidth with N2796A: 1 GHz

Replacement Kit Part Numbers (the quantities are the amounts shipped in the replacement kits, not the quantity included with the probe):


Y-Lead Adapter: 5061-7392 (qty. 1)

The Y-lead adapter offers a convenient and reliable method to connect both the probe signal and ground to probe points on the board or other probe accessories. For example, this can be used to connect to connector headers or to other accessories like the Flex Nose Clip Adapter (see the pictures below).

The sockets accept 0.56-0.86 mm round and 0.64 mm square (0.022 - 0.034 inch round and 0.025 inch square) diameter pins. The maximum insertion depth is 7.5 mm (0.295 inch) and the outer insulation of the socket is < 2.54 mm. You can use it for 2.54 mm pitch (0.1 inch) terminals.



Configuration 5: Flex Nose Clips



Bandwidth with N2795A: 500 MHz

Bandwidth with N2796A: 500 MHz

Replacement Kit Part Numbers (the quantities are the amounts shipped in the replacement kits, not the quantity included with the probe):

Y-Lead Adapter: 5061-7392 (qty. 1)

Flex Nose Clip (red): 5061-7390 (qty. 1)

Flex Nose Clip (black): 5061-7391 (qty. 1)

You can use flex nose clips to make connections to components or wires with leads that are 1.01 mm (0.04 inches) in diameter or smaller. With today's miniature IC- and component-packaging techniques, these clips can make probing challenging devices much easier. Most users prefer to attach the clips to the probe via the Y-lead adapter as shown in the picture below. Press the back of the flex nose clip to extend the grasping hook and then release to tighten the hook around the component you are testing.



Configurations 6 and 7: Ground Leads and Right Angle Ground Leads



Replacement Kit Part Numbers (the quantities are the amounts shipped in the replacement kits, not the quantity included with the probe):

6 cm Ground Lead: 5061-7395 (qty. 2)

12 cm Ground Lead: 5061-7396 (qty. 2)

5 cm Right Angle Ground Lead: 5061-7399 (qty. 2)

10 cm Right Angle Ground Lead: 5061-7400 (qty. 2)

These ground leads can be used to reach grounding locations that are farther away from the probing location than can be reached by either the ground blade or offset ground. However, the longer leads means they have a larger inductance in the ground return path which corresponds to a lower performance than the ground blade and offset ground.

Please note that the straight leads are not included with the probe. They must be ordered as a replacement kit. For more information about the replacement kits, see page 6.

Color Coding Rings

The color coding rings that can be used to keep track of which probe is connected to which channel input on your oscilloscope. Place one ring on the probe cable near the oscilloscope input and place another ring of the same color near the probe head. This ensures that you can pick up a probe and immediately know which channel it is connected to without having to track the cable back to the oscilloscope channel input.

Calibration Testing Procedures

These procedures are used to test the warranted specifications for the N2795A/N2796A single-ended active probes. The recommended calibration test interval for the N2795A/N2796A is once a year or as required. Use the equipment listed in the “Test Equipment Required” section to complete the Calibration Testing Procedures.

To Test Input Resistance

Required Test Equipment

Description	Minimum Requirements	Part Number
Digital Multimeter (DMM)	resistance $\pm 1\%$	Agilent 34401A

Test Probe Input Resistance

- 1 Connect the DMM probes between the probe tip and ground at the tip of the probe.
 - 2 Set up the DMM to measure resistance. The resistance should read $1\text{ M}\Omega + 0\%$, -2.5% .
 - 3 Record the resistance.
-

To Test Bandwidth

This test ensures that the Probe meets its specified bandwidth.

N2795A: 1 GHz

N2796A: 2 GHz

Equipment/Tool	Critical Specification	Model Number
Vector Network Analyzer (VNA)	13 GHz sweep range full 2 port cal Option 1D5	Agilent 8720ES
Calibration Standards	No Substitute	Agilent 85052D
External Power Supply	No Substitute	Agilent 1143A
AutoProbe Interface Adapter	No Substitute	Agilent N1022A
Outside thread 3.5 mm (male) to 3.5 mm (female) adapter	No Substitute	Agilent 5062-1247
Cable (2)	3.5 mil; SMA; High Quality	Agilent 8120-4948
Cable	1.5 mil Probe Power Extension No Substitute	Agilent 01143-61602
PV/DS Test Board	No Substitute (In E2655B Kit)	Agilent E2655-66503


Using the 8720ES VNA successfully

Remember these simple guidelines when working with the 8720ES VAN to get accurate stable measurements.

- 1 Sometimes it may take a few seconds for the waveforms to settle completely. Please allow time for waveforms to settle before continuing.
- 2 Make sure all connections are tight and secure. If needed, use a vise to hold the cables and test board stable while making measurements.
- 3 Be careful not to cross thread or force any connectors. This could be a very costly error to correct.

Initial Setup

- 1 Turn on the 8720ES VNA and let warm up for 20 minutes.
- 2 Press the green "Preset" key on the 8720ES VNA.
- 3 Use the 8720ES VNA's default power setting of 0 dBm. You can locate this feature by pressing the "Power" key on the front panel.
- 4 Set the 8720ES VNA's averaging to 4. You can find this selection menu by pressing the "AVG" key. Then select the "Averaging Factor" screen key to adjust the averaging.
- 5 Press the "Sweep Setup" key on the 8720ES VNA. Then press the "sweep type menu" screen key. Select the "log freq" screen key.
- 6 Connect the probe under test to the Auto Probe Adapter and power the probe using the 1143A power supply. Install the outside thread adapter to the Auto Probe Adapter.




01191E12

To Test Bandwidth


Calibrating a Reference Plane

To get a reliable measurement from the 8720ES VNA we must calibrate a reference plane so that the 8720ES VNA knows where the probe under test is located along the transmission line.

- 1 Press the "Cal" key on the 8720ES VNA.




- 2 Then Press the "cal menu" screen key.
- 3 Finally, press the "full 2 port" screen key.
- 4 Connect one of the high quality SMA cables to port one and to the pincher side of PV/DS test board.
- 5 The calibration reference plane is at the other end of PV/DS test board.



To Test Bandwidth

- 6 Perform Calibration for the port one side of the Reference plane.
 - Press the "reflection" screen key
 - Connect open end of 85052D to the non-pincher side of the PV/DS test board.
 - Select the "open" screen key under the "Forward" group.
 - The 8720ES VAN will beep when done.
 - Connect short end of 85052D to the non-pincher side of the PV/DS test board.
 - Select "short" screen key under the "Forward" group.
 - The 8720ES VAN will beep when done.
 - Connect load end of 85052D to the non-pincher side of the PV/DS test board.
 - Select the "loads" screen key under the "Forward" group.
 - Press "broadband" screen key selection.
 - The 8720ES VAN will beep when done.
 - Press the "done loads" screen key.
 - You have just calibrated one side of the reference plane.

7 Connect the other high quality SMA cable to port two of the 8720ES VNA.




8 Get the opposite sex of the 85052D calibration standards for the next step.

9 Perform Calibration for the port two side of the Reference plane.

- Press the "reflection" screen key.
- Connect open end of 85052D to the available end of the port two SMA cable.
- Select "open" screen key under the "Reverse" group.
- The 8720ES VNA will beep when done.
- Connect short end of 85052D to the available end of the port two SMA cable.
- Select "short" screen key under the "Reverse" group.

To Test Bandwidth

- The 8720ES VNA will beep when done.
 - Connect load end of 85052D to the available end of the port two SMA cable.
 - Select the "loads" screen key the "Reverse" group.
 - Press "broadband" screen key selection.
 - The 8720ES VNA will beep when done.
 - Press the "done loads" screen key.
 - You have just calibrated the other side of the reference plane.
- 10** Press "standards done" key.
- 11** Connect port two SMA cable to the non-pincher side of PV/DS test board.



- 12 Press the "transmission" screen key.
- 13 Press the "do both fwd and reverse" screen key.
- 14 The 8720ES VNA will beep **four** times when done.
- 15 Press the "isolation" screen key.
- 16 Press the "omit isolation" screen key.
- 17 Press "done 2 port cal" screen key.
- 18 Set the 8720ES VNA's averaging to off.
- 19 Save the reference plane cal by pressing the "save recall" key then the "save state" key.
- 20 You may change name if you wish.
- 21 Press the "scale reference" key. Then
 - Set for 1 dB per division.
 - Set reference position for 7 divisions.
 - Set reference value for 0 dB
- 22 Press the "measure" key.
- 23 Press the "s21" screen key.
- 24 Ensure s21 response on screen is flat out to 1 GHz (N2795A) or 2 GHz (N2796A).

Measuring Vin Response

- 1 Position the probe conveniently to allow the probe tip to be normal to the PV/DS board.
- 2 Hold the probe in position, or use a positioner, so that the signal and ground are making contact and are perpendicular to the fixture.
Press the "Sweep Setup" key on the 8720ES VNA. Then press the "trigger menu" screen key. Select the "continuous" screen key.
- 3 You should now have the Vin waveform on screen.
- 4 Select "display key" then "data->memory" screen key.
- 5 You have now saved Vin waveform into the 8720ES VNA's memory for future use.

Measuring Vout Response

- 1 Disconnect the port 2 cable from PV/DS test board and attach to probe output on the AutoProbe Adapter.
- 2 Connect the 85052D cal standard load to PV/DS test board (non-pincher side).
- 3 Press "scale reference" key on the 8720ES VNA.
- 4 Set reference value to -20 dB.
- 5 Hold probe in place as described previously.
- 6 The display on screen is Vout.

Displaying Vout/Vin Response on 8720ES VNA Screen

- 1 Press the "Display" Key.
- 2 Then select the "Data/Memory" Screen Key. You may need to adjust the "Reference Value", located under the "Scale Ref" key, slightly to position the waveform at center screen.
- 3 Press marker key and position the marker to the first point that the signal is -3 dB below center screen.
- 4 Read marker frequency measurement and record it in the test record located later in this chapter.
- 5 The bandwidth test passes if the frequency measurement is greater than the probe's bandwidth limit. Example: > 1 GHz (N2795A) or 2 GHz (N2796A).

To Return the Probe to Agilent Technologies for Service

If the probe is found to be defective we recommend sending it to an authorized service center for all repair and calibration needs. Follow the following steps before shipping the probe back to Agilent Technologies for service.

- 1 Contact your nearest Agilent sales office for information on obtaining an RMA number and return address.
- 2 Write the following information on a tag and attach it to the malfunctioning equipment.
 - Name and address of owner
 - Product model number Example N2796A
 - Product Serial Number Example MYXXXXXXXXX
 - Description of failure or service required

Include probing and browsing tips if you feel the probe is not meeting performance specifications or a yearly calibration is requested.

- 3 Protect the Probe by wrapping in plastic or heavy paper.
- 4 Pack the Probe in the original carrying case or if not available use bubble wrap or packing peanuts.
- 5 Place securely in sealed shipping container and mark container as "FRAGILE".

If any correspondence is required, refer to the product by serial number and model number.

Safety Notices

This apparatus has been designed and tested in accordance with IEC Publication 1010, Safety Requirements for Measuring Apparatus, and has been supplied in a safe condition. This is a Safety Class I instrument (provided with terminal for protective earthing). Before applying power, verify that the correct safety precautions are taken (see the following warnings). In addition, note the external markings on the instrument that are described under "Safety Symbols."

Warnings

- Whenever it is likely that the ground protection is impaired, you must make the instrument inoperative and secure it against any unintended operation.
- Service instructions are for trained service personnel. To avoid dangerous electric shock, do not perform any service unless qualified to do so. Do not attempt internal service or adjustment unless another person, capable of rendering first aid and resuscitation, is present.
- Do not install substitute parts or perform any unauthorized modification to the instrument.
- Capacitors inside the instrument may retain a charge even if the instrument is disconnected from its source of supply.
- Do not operate the instrument in the presence of flammable gasses or fumes. Operation of any electrical instrument in such an environment constitutes a definite safety hazard.
- Do not use the instrument in a manner not specified by the manufacturer.

Safety Symbols



Instruction manual symbol: the product is marked with this symbol when it is necessary for you to refer to the instruction manual in order to protect against damage to the product or personal injury.



Hazardous voltage symbol.



Earth terminal symbol: Used to indicate a circuit common connected to grounded chassis.

Notices

© Agilent Technologies, Inc.
2010

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number
N2795-97000, October 2010

Print History
N2795-97000, October 2010

Agilent Technologies, Inc.
1900 Garden of the Gods Road
Colorado Springs, CO
80907 USA

Restricted Rights Legend
If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as "Commercial computer software" as defined in DFAR 252.227-7014 (June 1995), or as a "commercial item" as defined in FAR 2.101(a) or as "Restricted computer software" as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies' standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Document Warranty

The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the

maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

Agilent Technologies
P.O. Box 2197
1900 Garden of the Gods Road
Colorado Springs, CO 80901

www.valuetronics.com



Agilent Technologies

Agilent Technologies, October 2010
Manual Part Number: N2795-97000
Only Available Electronically

