1413 Precision

# **Decade Capacitor**

**User and Service Manual** 



Copyright © 2003 IET Labs, Inc.

1413 im/May, 2003



# Contents

WARRANTY
WARNING v
CAUTIONv
Section 1
INTRODUCTION1
1.1 Purpose
1.2 Description
1.3 Controls and Connectors 1
1.4 Inspection and Unpacking2
Section 2
SPECIFICATIONS
Section 3
OPERATION PROCEDURE4
3.1 Dimensions
3.2 Mounting
3.3 Connections
3.3.1 Three-Terminal Capacitor
Section 4
THEORY6
4.1 Circuit Description
4.2 Frequency Characteristics
Section 5
SERVICE
5.1 Customer Service7
5.2 Instrument Return
Section 6
MAINTENANCE8
6.1 Minimum Performance Standards
6.1.1 General
6.1.2 Capacitance
6.1.3 Terminal Capacitance
6.1.4 Insulation Resistance
6.2 Adjustments 11
6.3 Knob Removal and Replacement 11
Section 7
PARTS LISTS AND DIAGRAMS

# Figures

Figure 1-1. Elementary schematic diagram of the Type 1413 Decade
Capacitor. Connections for all 24 capacitors are similar to those
for the three capacitors shown 1
Figure 1-2. Front panel on the Capacitor.
Six 11-position switches with knobs and dials to set desired value
of capacitance between HIGH and LOW terminals
Figure 1.3 Connection on the rear panel of the 1413
Figure 3-1. Dimensions of the bench and relay-rack models
Figure 6-1. Top interior view of the Capacitor
Figure 7-2. Schematic circuit diagram for the 1413 Precision Decade Ca-
pacitor. Switches are shown in zero position

# WARRANTY

We warrant that this product is free from defects in material and workmanship and, when properly used, will perform in accordance with applicable IET specifications. If within one year after original shipment, it is found not to meet this standard, it will be repaired or, at the option of IET, replaced at no charge when returned to IET. Changes in this product not approved by IET or application of voltages or currents greater than those allowed by the specifications shall void this warranty. IET shall not be liable for any indirect, special, or consequential damages, even if notice has been given to the possibility of such damages.

THIS WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUD-ING BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OF MERCHANTIBILITY OR FITNESS FOR ANY PARTICULAR PURPOSE.



# OBSERVE ALL SAFETY RULES WHEN WORKING WITH HIGH VOLTAGES OR LINE VOLTAGES.

Dangerous voltages may be present inside this instrument. Do not open the case Refer servicing to qulified personnel

HIGH VOLTAGES MAY BE PRESENT AT THE TERMINALS OF THIS INSTRUMENT

WHENEVER HAZARDOUS VOLTAGES (> 45 V) ARE USED, TAKE ALL MEASURES TO AVOID ACCIDENTAL CONTACT WITH ANY LIVE COMPONENTS.

USE MAXIMUM INSULATION AND MINIMIZE THE USE OF BARE CONDUCTORS WHEN USING THIS INSTRUMENT.

Use extreme caution when working with bare conductors or bus bars.

WHEN WORKING WITH HIGH VOLTAGES, POST WARNING SIGNS AND KEEP UNREQUIRED PERSONNEL SAFELY AWAY.



DO NOT APPLY ANY VOLTAGES OR CURRENTS TO THE TERMINALS OF THIS INSTRUMENT IN EXCESS OF THE MAXIMUM LIMITS INDICATED ON THE FRONT PANEL OR THE OPERATING GUIDE LABEL.

## Section 1 Introduction

#### 1.1 Purpose

The Type 1413 Precision Decade Capacitor is both a high-quality, wide-range standard and a reliable component for systems. It can be used as a bench model for versatility, or it can be rack-mounted.

This six-decade capacitor features fine adjustment over a wide range of capacitance, with excellent accuracy. Any value in the range of 0 to 1.111 11  $\mu$ F can be set, with a resolution of 1 pF.

Other units up to 10,000 µF are available; consult IET.

#### **1.2 Description**

The six decades have steps of 1, 10, 100, 1000 pF, and 0.01 and 0.1  $\Box$ F. Air capacitors are used for the two lower decades, with precision silvered-mica capacitors for the others. Air trimmers are used for trimming the two lowest silvered-mica decades.

The connections to the inner shield and to the case are shown in Figure 1-1. Low terminal-to-guard and detector-input capacitances are obtained by dividing the shielding into two parts. When the two parts are connected together, the 1413 becomes a well-shielded three-terminal capacitor with extremely low zero capacitance. The connections are made by means of BNC connectors on the rear of the cabinet. The inner contacts provide the connections to the capacitors. The outer shells connect to the shields and, when connected together, form the third terminal of the capacitor.

The stability of the 1413 (refer to Specifications) is such that it should not require readjustment with normal service. However, should it become desirable, the four lower decades contain trimmer capacitors that are accessible for this purpose (refer to Section 5.4).



Figure 1-1. Elementary schematic diagram of the Type 1413 Decade Capacitor. Connections for all 24 capacitors are similar to those for the three capacitors shown.

#### **1.3 Controls and Connectors**

The controls and connectors on the front and rear of the 1413 are shown in Figures 1-2 and 1-3.

### 1.4 Inspection and Unpacking

Inspection - If the shipping carton appears damaged, contact the carrier immediately and request that their agent be present when the instrument is unpacked. If the instrument appears to be damaged when unpacked, have the agent witness its condition, take photos of the instrument and carton, and retain the carton until the matter is resolved. Perform an electrical check immediately to determine if any internal damage has occurred and file any required claims with the carrier. It may be necessary to return the instrument to IET

for evaluation, repair estimate and actual repair.

Unpacking - The instrument is wrapped in plastic, and desiccant may be included if the environmental conditions and destination warrant. Do not unwrap the instrument until ready to use or install. For bench use, the bail will raise the front of the instrument if desired. Refer to 2.2 for rack-mounting instructions.



Figure 1-2. Front panel on the Capacitor. Six 11-position switches with knobs and dials to set desired value of capacitance between HIGH and LOW terminals.



Figure 1-3. Connection on the rear panel of the 1413.

# Section 2 Specifications

**Range:** 0 to 1.111 11  $\Box$ F. controlled by six in-line-readout dials.

Accuracy: ±(0.05% + 0.5 pF) at 1 kHz.

Stability: ±(0.01% +0.1 pF) per year. Temperature coefficient: <<20 ppm/°C from 10 to 50°C.

Zero Capacitance: <0.1 pF.

Voltage Rating: 500 V pk max up to 10 kHz.

	1 pF to 100 pF	101 pF to 1000 pF	1001 pF to 2000 pF	2001 pF to 0.1 μF	0.1 μF to 1.11111 μF
Dissipation Factor, max at 1 kHz	0.002	0.001	0.0005	0.0003	0.0004
Insulation Resistance, 3 term., after 2 m at 500 V dc	≥ 5 x 10 <sup>10</sup> Ω			≥ 5 x 10° Ω	
Terminal Capa- citance, max high to case high to guard low to guard	4 pF 80 pF 45 pF	8 pF 105 pF 70 pF	10 pF 115 pF 80 pF	30 pF 165 pF 110 pF	60 pF 200 pF 120 pF

**Interface:** Connections, 2 rear-mounted bnc connectors, adaptable to most other types.

**Available:** 0480-9703 Rack Adaptor Set to convert bench models to rack models.

**Mechanical:** Convertible bench/rack cabinet. Dimensions (w x h x d): 17 x 5.59 x 11.96 in. (432 x 142 x 304 mm); rack 19 x 5.22 x 10.9 in. (483 x 133 x 505 mm).

Weight: Bench 23 lb (10.5 kg) net, 29 lb (13.5 kg) shipping; rack 24 lb (11 kg) net, 30 lb (14 kg) ship-



Catalog Number	Description		
1413-9700	Bench Model		
1413-9701	Rack Model		
0480-9703	Rack-Adaptor Set		

# Section 3 Operating Procedure

mounting (P/N 1413-9701), it is shipped with Rack-Adaptor Set P/N 0480-9703, which includes all hardware necessary for mounting the instrument in a stan-

dard 19-in. rack or cabinet (refer to Table 3-1). (Or-

der the Rack-Adaptor set alone if a bench model is to

be converted for rack mounting.)

### 3.1 Dimensions

The dimensions of both the bench and relay-rack models of the 1413 are shown in Figure 3-1, and are also listed in Section 2.

### 3.2 Mounting

The 1413 may be ordered enclosed in a cabinet for bench use (P/N 1413-9700). If ordered for rack



Figure 3-1. Dimensions of the bench and relay-rack models.

Table 3-1PARTS INCLUDED IN RACK-ADAPTOR SETP/N 0480-9703			
Fig. 3-2	No.		
Ref	Used	Description	Part No.
С	2	Rack-Adaptor Assembly (handle)	0480-4903
	1	Hardware Set	0480-3080
		Includes:	
D		4 Screws, BH 10-32x 5/16"	
Е		4 Screws, BH 10-32x9/16", with nylon cup washers	

#### CAUTION

Remove the cabinet only when converting the instrument from bench to rack mounting or vice versa. Use extreme care. Do not change the position of the leads or components, as doing so may alter the calibration. IET recommends taking accurate spot readings before and after this procedure in order to verify that the instrument's electrical characteristics have not been affected. Cover the unit with a plastic sheet when working on the cabinet. Clean the inside of the cabinet thoroughly when installation is complete. Replace the instrument in the cabinet as soon as possible after the work is completed.

To install the instrument in a rack:

- a. Loosen the four captive 10-32 screws in the rear of the cabinet until the chassis is free; slide the chassis forward, out of the cabinet.
- b. Match drill the cabinet for proper positioning of the rack-mount adaptors.
- c. Remove the four feet and the bail from the cabinet if required for clearance.
- d. Attach one Rack Adaptor Assembly (G) to each side of the cabinet using the hardware supplied.
- e. Install the instrument in the cabinet and lock it in place with the four captive screws in the rear panel that were loosened in step a.

f. Slide the entire assembly into the relay rack and install it with the four 9/16-inch screws with captive nylon cup washers (E). Use two screws on each side and tighten them by inserting a screwdriver through the holes in the handles.

To reconvert the instrument for bench use, reverse the above procedure. Seal the mounting holes if the working environment warrants such action.

#### **3.3 Connections**

#### CAUTION

The voltage rating of the 1413 is 500 V pk maximum up to 10 kHz, varying approximately inversely with frequency above 10 kHz. Do not exceed these values.

#### 3.3.1 Three-Terminal Capacitor

When the 1413 is used as a three-terminal capacitor, the outer conductors (shells) of the two bnc connectors must be connected together, to provide complete shielding.

If binding-post connections are desired, these can be provided by standard adaptors plugged into the bnc connectors on the rear of the 1413. Plug the adaptors into the 1413 with the shield binding posts adjacent to each other. These shield posts can then be connected together with a shorting link.

# Section 4 Theory

### **4.1 Circuit Description**

Each decade consists of four capacitors with values in the ratio of 1-2-2-5. They are connected in parallel in different combinations to provide ten equal steps of capacitance. Individual capacitors that are not used at a particular setting of the switches are completely disconnected from the circuit and their high side is connected to the inner or "guard" shield. The inner shield and the wiring are arranged to keep the capacitance at the zero setting to a very low value (< 0.1 pF) and to keep the capacitance from HIGH to case and from HIGH and LOW to the inner shield to a minimum. An elementary schematic is shown in Figure 1-1 and the complete schematic circuit diagram is given in Figure 7-1.

### **4.2 Frequency Characteristics**

Typical variations of capacitance and dissipation factor with frequency are shown in Section 2.

## Section 5 Service

### **5.1 Customer Service**

The IET warranty attests to the quality of materials and workmanship in our products. For application assistance or if difficulties occur, our engineers will assist in any way possible. If you cannot eliminate the difficulty, please e-mail, FAX, or phone our Service Department, giving full information of the trouble and of steps taken to remedy it. Be sure to include the type and serial number of the instrument.

In the US call:

800-475-1220 or 617-969-0804 for technical support

800-899-8438 or 516-334-5959 for customer service

516-334-5988 for FAX

www.ietlabs.com

#### **5.2 Instrument Return**

Before returning an instrument to IET for service please call our Service Department at 800-899-8438 for a Return Material Authorization (RMA). Supply a Purchase Order Number or Credit Card information to insure expedient processing. Units under warranty will be repaired at no charge. For any questions on repair costs or shipping instructions, please contact our Service Department at the above number. To safeguard an instrument during shipment, please use packaging that is adequate to protect it from damage, (i.e., equivalent to the original packaging) and mark the box "Delicate Electronic Instrument" and also with the RMA number. Material should be sent freight prepaid to:

> IET Labs, Inc. 10 Dedham Street Newton Highlands, MA 02461

Attention: Service Department

	Table 6-2		
Connections for Terminal-Capacitance Measurements			
Measurement	Connection	Lead Type	
HIGH to Case	Bridge L to 1413 HIGH	Patch cord	
	Bridge H to outer shell of 1413 LOW	Shielded clip lead	
	Bridge GND to inner conductor of	Single banana plug	
	1413 LOW	patch cord	
HIGH to Guard	Bridge H to shell, and L to inner conductor of 1413 HIGH	Shielded clip leads	
	Bridge GND to shell and inner	Patch cord with	
	conductor of 1413 LOW	alligator clips	
LOW to Guard	Bridge L to 1413 LOW Bridge H to shell of 1413 HIGH Bridge GND to inner conductor of 1413 HIGH	Patch cord Shielded clip lead Single banana plug patch cord	



Figure 6-1. Top interior view of the Capacitor.

### ELECTRICAL PARTS LIST

### Ref Des

Description

Part No.

IET

## CAPACITORS

	A: D: 1 4: 1450 E	1200 2500
	Air Dielectric, 1.4-5.0 pF	4380-3500
	Air Dielectric, 1.4-5.0 pF	4380-3500
	Air Dielectric, 1.4-5.0 pF	4380-3500
C4	Air Dielectric, 1.7-8.7 pF	4380-3600
C5	Quartz, Var., 0.6-1.8 pF,	4910-1180
Cll	Air Dielectric, 2.7-19.6 pF	4380-3700
C12	Air Dielectric, 2.7-19.6 pF	4380-3700
C13	Air Dielectric, 2.7-19.6 pF	4380-3700
C14	Air Dielectric, 2.7-19.6 pF	4380-3700
C15	Ceramic, 33 pF ±5%, NPO, 500V	4410-0335
C21	96.3 pF ±1 pF	0505-4030
C22	196.0 pF ±1 pF	0505-4031
C23	$196.0 \text{ pF} \pm 1 \text{ pF}$	0505-4031
C24	496.0 pF ±1.3 pF	0505-4032
C25	Air Dielectric, 1.7-8.7 pF	4380-3600
C26	Air Dielectric, 1.7-8.7 pF	4380-3600
C27	Air Dielectric, 1.7-8.7 pF	4380-3600
C28	Air Dielectric, 1.7-8.7 pF	4380-3600
C31	986.1 pF ±2.1 pF	0505-4033
C32	$1989  \text{pF} \pm 5  \text{pF}$	0505-4034
C33	$1989 \text{ pF} \pm 5 \text{ pF}$	0505-4034
C34	$4989  \text{pF} \pm 5  \text{pF}$	0505-4035
C35	Air Dielectric, 2.7-19.6 pF	4380-3700
C36	Air Dielectric, 2.7-19.6 pF	4380-3700
C37	Air Dielectric, 2.7-19.6 pF	4380-3700
C38	Air Dielectric, 2.7-19.6 pF	4380-3700
C41	$10,000 \text{ pF} \pm 2.5 \text{ pF}$	0505-4036
C42	$20,000 \text{ pF} \pm 5 \text{ pF}$	0505-4037
C43	$20,000 \text{ pF} \pm 5 \text{ pF}$	0505-4037
C44	$.05 \mu\text{F} \pm 12.5 \text{pF}$	0505-4038
C51	$0.1 \mu\text{F} \pm 12.5 \text{pF}$	0505-4039
C52	$0.2 \mu\text{F} \pm 50 \text{pF}$	0505-4040
C53	$0.2 \mu\text{F} \pm 50 \text{pF}$	0505-4040
C54	$0.5 \mu\text{F} \pm 125 \mu\text{F}$	0505-4041
	• - <b>F</b>	

MECHANICAL PARTS LIST					
	Front View (See Fig 1-2)				
			IET		
Qty	Ref	Description	Part No.		
			44.04.0001		
1	I	Cabinet asm., includes: Bail*	4181-2901		
1		Gasket (mat'l.)	0051-0006		
1		Foot, left front*			
1		Foot, right front*			
2		*Poot, 1 & r, rear* *Part of foot/bail kit			
6	2	Knob/Dial assembly	350020		
6	3	Cap, knob	350019		
		1 /			
Rear View (See Fig 1-3)					
			IFT		
Qty	Ref	Description	Part No.		
1	1	Connector, bnc, isolated (Std.) (Other types optional)	31-10/800-2540-03		
1	2	Connector, bnc (Other types optional)	31-221/UG-1094		
		(outer types opticituit)			