

Keysight Technologies  
85038A/F/M 7-16  
Calibration Kits



NOTICE: This document contains references to Agilent Technologies. Agilent's former Test and Measurement business has become Keysight Technologies. For more information, go to [www.keysight.com](http://www.keysight.com).



---

## **WARRANTY STATEMENT**

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, KEYSIGHT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. KEYSIGHT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR ANY INFORMATION CONTAINED HEREIN. SHOULD KEYSIGHT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS DOCUMENT THAT CONFLICT WITH THESE TERMS, THE WARRANTY TERMS IN THE SEPARATE AGREEMENT WILL CONTROL.

---

## **DFARS/Restricted Rights Notice**

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as "Commercial computer software" as defined in DFAR 252.227-7014 (June 1995), or as a "commercial item" as defined in FAR 2.101(a) or as "Restricted computer software" as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Keysight Technologies' standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

---

## **Certification**

Keysight Technologies certifies that this product met its published specifications at the time of shipment from the factory. Keysight Technologies further certifies that its calibration measurements are traceable to the United States National Institute of Standards and Technology, to the extent allowed by the Institute's calibration facility, and to the calibration facilities of other International Standards Organization members.

---

## Assistance

Product maintenance agreements and other customer assistance agreements are available for Keysight products.

For any assistance, contact Keysight Technologies. Refer to "[Contacting Keysight](#)," on page 5-5.

---

## Printing Copies of Documentation from the Web

To print copies of documentation from the Web, download the PDF file from the Keysight web site:

- Go to [www.keysight.com](http://www.keysight.com).
- Enter the document's part number (located on the title page) in the **Search** box.
- Click Search.
- Click on the hyperlink for the document.
- Click the printer icon located in the tool bar.



## 1 General Information

Calibration Kit Overview . . . . .	1-2
Kit Contents . . . . .	1-2
Calibration Definitions . . . . .	1-2
Equipment Required but Not Supplied . . . . .	1-2
Incoming Inspection . . . . .	1-3
Serial Numbers . . . . .	1-4
Recording the Device Serial Numbers . . . . .	1-4
Clarifying the Terminology of a Connector Interface . . . . .	1-5
Preventive Maintenance . . . . .	1-5
When to Calibrate . . . . .	1-6

## 2 Specifications

Environmental Requirements . . . . .	2-2
Temperature—What to Watch Out For . . . . .	2-2
Mechanical Characteristics . . . . .	2-3
Pin Depth . . . . .	2-3
Electrical Specifications . . . . .	2-4
Certification . . . . .	2-5
Connector Standards . . . . .	2-5

## 3 Use, Maintenance, and Care of the Devices

Electrostatic Discharge . . . . .	3-2
Visual Inspection . . . . .	3-3
Look for Obvious Defects and Damage First . . . . .	3-3
Inspect the Mating Plane Surfaces . . . . .	3-3
Inspect Female Connectors . . . . .	3-3
Cleaning Connectors . . . . .	3-4
Connections . . . . .	3-5
How to Make a Connection . . . . .	3-5
Final Connection Using a Torque Wrench . . . . .	3-5
How to Separate a Connection . . . . .	3-8
Handling and Storage . . . . .	3-8

## 4 Performance Verification

Introduction . . . . .	4-2
How Keysight Verifies the Devices in Your Kit . . . . .	4-2
Recertification . . . . .	4-3
How Often to Recertify . . . . .	4-3
Where to Send a Kit for Recertification . . . . .	4-3

## 5 Troubleshooting

Troubleshooting Process . . . . .	5-2
Where to Look for More Information . . . . .	5-3
Returning a Kit or Device to Keysight . . . . .	5-4
Contacting Keysight . . . . .	5-5

## 6 Replaceable Parts

---

# Contents

Introduction .....	6-2
<b>A Standard Definitions</b>	
Class Assignments and Standard Definitions Values are Available on the Web .....	A-2

# 1 General Information

## Calibration Kit Overview

The Keysight 85038 7-16 calibration kits are used to calibrate Keysight network analyzers up to 7.5 GHz for measurements of components with 7-16 connectors.

### Kit Contents

Use the Contents List in the shipping container to verify the completeness of your shipment. Although this list is the most accurate, you can also use the illustration in Chapter 7 to verify the items in your shipment. If your shipment is not complete, contact Keysight Technologies - refer to "[Contacting Keysight](#)," on page 5-5.

Refer to [Table 6-1](#) for a list of kit contents part numbers.

### Calibration Definitions

The calibration kit must be selected and the calibration definitions for the devices in the kit installed in the network analyzer prior to performing a calibration. Refer to your network analyzer user's guide for instructions on selecting the calibration kit and performing a calibration.

The calibration definitions can be:

- resident within the analyzer
- entered from the front panel

Class assignments and standard definitions may change as more accurate model and calibration methods are developed. You can download the most recent class assignments and standard definitions from Keysight's Calibration Kit Definitions Web page at

<http://na.support.keysight.com/pna/caldefs/stddefs.html>

### Installation of the Calibration Definitions

The calibration definitions for the kit may be permanently installed in the internal memory or hard disk of the network analyzer.

If the calibration definitions for the kit are not permanently installed in the network analyzer, they must be manually entered. Refer to your network analyzer user's guide for instructions.

### Equipment Required but Not Supplied

Connector cleaning supplies and various electrostatic discharge (ESD) protection devices are not supplied with the calibration kit but are required to ensure successful operation of the kit. Refer to [Table 6-2 on page 6-5](#) for ordering information.

## Incoming Inspection

Check for damage. The foam-lined storage case provides protection during shipping. Verify that this case and its contents are not damaged.

If the case or any device appears damaged, or if the shipment is incomplete, contact Keysight Technologies. For contact information, see “[Contacting Keysight](#)” on page 5-5. Keysight will arrange for repair or replacement of incomplete or damaged shipments without waiting for a settlement from the transportation company.

When you send the kit or device to Keysight, include a service tag (found near the end of this manual) with the following information:

- your company name and address
- the name of a technical contact person within your company, and the person's complete phone number
- the model number and serial number of the kit
- the part number and serial number of the device
- the type of service required
- a *detailed* description of the problem

## Serial Numbers

A serial number is attached to this calibration kit. The first four digits followed by a letter comprise the serial number prefix; the last five digits are the suffix, unique to each calibration kit.

### Recording the Device Serial Numbers

In addition to the kit serial number, the devices in the kit are individually serialized (serial numbers are labeled onto the body of each device). Record these serial numbers in **Table 1-1**. Recording the serial numbers will prevent confusing the devices in this kit with similar devices from other kits.

**Table 1-1 Serial Number Record for the 85038A/F/M**

Device	Serial Number
<b>Calibration kit</b>	_____
Open -m-	_____
Open -f-	_____
Short -m-	_____
Short -f-	_____
Load -m-	_____
Load -f-	_____
<b>Adapters (7-16)</b>	_____
-m- to -m-	_____
-f- to -f-	_____

---

## Clarifying the Terminology of a Connector Interface

In this document and in the prompts of the PNA calibration wizard, the sex of cable connectors and adapters is referred to in terms of the center conductor. For example, a connector or device designated as 1.85 mm –f– has a 1.85 mm female center conductor.

**8510-series, 872x, and 875x ONLY:** In contrast, during a measurement calibration, the network analyzer softkey menus label a 1.85 mm calibration device with reference to the sex of the analyzer’s test port connector—not the calibration device connector. For example, the label SHORT(F) refers to the short that is to be connected to the female test port. This will be a male short from the calibration kit.

**Table 1-2 Clarifying the Terminology of Connectors: Examples**

Terminology	Meaning
Short –f–	Female short (female center conductor)
Short (f)	Male short (male center conductor) to be connected to female port

A connector gage is referred to in terms of the connector that it measures. For instance, a male connector gage has a female connector on the gage so that it can measure male devices.

---

## Preventive Maintenance

The best techniques for maintaining the integrity of the devices in your kit include:

- routine visual inspection
- cleaning
- proper connection techniques

All of these are described in [Chapter 3, “Use, Maintenance, and Care of the Devices.”](#) Failure to detect and remove dirt or metallic particles on a mating plane surface can degrade repeatability and accuracy and can damage any connector mated to it. Improper connections, resulting from pin depth values being out of limit (see [Table 2-2 on page 2-4](#)) or from bad connection techniques, can also damage these devices.

---

## When to Calibrate

A network analyzer calibration remains valid as long as the changes in the systematic error are insignificant. This means that changes to the uncorrected leakages (directivity and isolation), mismatches (source match and load match), and frequency response of the system are small (<10%) relative to

**When to Calibrate**

accuracy specifications.

Change in the environment (especially temperature) between calibration and measurement is the major cause in calibration accuracy degradation. The major effect is a change in the physical length of external and internal cables. Other important causes are dirty and damaged test port connectors and calibration standards. If the connectors become dirty or damaged, measurement repeatability and accuracy is affected. Fortunately, it is relatively easy to evaluate the general validity of the calibration. To test repeatability, remeasure one of the calibration standards. If you can not obtain repeatable measurements from your calibration standards, maintenance needs to be performed on the test port connectors, cables and calibration standards. Also, maintain at least one sample of the device under test or some known device as your reference device. A verification kit may be used for this purpose. After calibration, measure the reference device and note its responses. Periodically remeasure the device and note any changes in its corrected response which can be attributed to the test system. With experience you will be able to see changes in the reference responses that indicate a need to perform the measurement calibration again.

## **2 Specifications**

## Environmental Requirements

**Table 2-1 Environmental Requirements**

Parameter	Limits
Operating temperature <sup>a</sup>	+15 °C to +35 °C (+59 °F to +95 °F)
Error-corrected temperature range <sup>b</sup>	±1 °C of measurement calibration temperature
Storage temperature	–40 °C to +75 °C (–40 °F to +167 °F)
Relative humidity	Type tested, 0% to 95% at 40 °C, non-condensing

- a. The temperature range over which the calibration standards maintain conformance to their specifications.
- b. The allowable network analyzer ambient temperature drift during measurement calibration and during measurements when the network analyzer error correction is turned on. Also, the range over which the network analyzer maintains its specified performance while correction is turned on.

### Temperature—What to Watch Out For

Changes in temperature can affect electrical characteristics. Therefore, the operating temperature is a critical factor in performance. During a measurement calibration, the temperature of the calibration devices must be stable and within the range shown in [Table 2-1](#).

---

**IMPORTANT** Avoid unnecessary handling of the devices during calibration because your fingers are a heat source.

---

---

## Mechanical Characteristics

Mechanical characteristics such as pin depth are *not* performance specifications. They are, however, important supplemental characteristics related to electrical performance.

### Pin Depth


Pin depth is the distance the center conductor mating plane differs from being flush with the outer conductor mating plane. See [Figure 2-1](#). Typical pin depth values for the devices in the kits are listed in [Table 2-2](#).

The pin depth of a connector can be in one of two states: either protruding or recessed.

**Protrusion** is the condition in which the center conductor extends beyond the outer conductor mating plane.

**Recession** is the condition in which the center conductor is set back from the outer conductor mating plane.

**Figure 2-1 Connector Pin Depth**



Pn42m

**Electrical Specifications****Table 2-2 Pin Depth Typical Values**

Device	Typical Pin Depth
	Inches Millimeters
Male devices	-0.068 to -0.069 -1.73 to -1.75
Female devices	-0.069 to -0.071 -1.75 to -1.80

**Electrical Specifications**

The electrical specifications in [Table 2-3](#) apply to the devices in your calibration kit when connected with a Keysight precision interface.

**Table 2-3 Electrical Specifications for 85038A/F/M Devices**

Device	Specification	Frequency (GHz)
Loads (male and female)	Return loss $\geq 40$ dB	dc to $\leq 4$
	Return loss $\geq 36$ dB	$> 4$ to $\leq 7.5$
Opens <sup>a</sup> (male and female)	$\pm 1^\circ$ deviation from nominal phase	dc to $\leq 4$
	$\pm 1.25^\circ$ deviation from nominal phase	$> 4$ to $\leq 7.5$
Shorts <sup>a</sup> (male and female)	$\pm 0.6^\circ$ deviation from nominal phase	dc to $\leq 4$
	$\pm 0.85^\circ$ deviation from nominal phase	$> 4$ to $\leq 7.5$
Adapters <sup>b</sup>	Return loss $\geq 38$ dB	dc to $\leq 7.5$

a. The specifications for the opens and shorts are given as allowed deviation from the nominal model as defined in the standard definitions (see [Appendix A](#), "Standard Definitions.").

b. Applies only to 85038F/M kits.

## **Certification**

Keysight Technologies certifies that this product met its published specifications at the time of shipment from the factory. Keysight further certifies that its calibration measurements are traceable to the United States National Institute of Standards and Technology (NIST) to the extent allowed by the institute's calibration facility, and to the calibration facilities of other International Standards Organization members. See "[How Keysight Verifies the Devices in Your Kit](#)," on page 4-2 for more information.

## **Connector Standards**

The 7-16 connectors in these calibration kits conform to or exceed the requirements for:

- IEC 169-4 reference grade
- DIN 47223
- EN 122190 grade 0 (European standard)
- BSEN 122190 (British standard)

Specifications  
**Electrical Specifications**

### **3 Use, Maintenance, and Care of the Devices**

**Electrostatic Discharge**

---

## Electrostatic Discharge


Protection against ESD (electrostatic discharge) is essential while connecting, inspecting, or cleaning connectors attached to a static-sensitive circuit (such as those found in test sets).

Static electricity can build up on your body and can easily damage sensitive internal circuit elements when discharged. Static discharges too small to be felt can cause permanent damage. Devices such as calibration components and devices under test (DUTs), can also carry an electrostatic charge. To prevent damage to the test set, components, and devices:

- *always* wear a grounded wrist strap having a 1 M $\Omega$  resistor in series with it when handling components and devices or when making connections to the test set.
- *always* use a grounded, conductive table mat while making connections.
- *always* wear a heel strap when working in an area with a conductive floor. If you are uncertain about the conductivity of your floor, wear a heel strap.
- *always* ground yourself before you clean, inspect, or make a connection to a static-sensitive device or test port. You can, for example, grasp the grounded outer shell of the test port or cable connector briefly.
- *always* ground the center conductor of a test cable before making a connection to the analyzer test port or other static-sensitive device. This can be done as follows:
  1. Connect a short (from your calibration kit) to one end of the cable to short the center conductor to the outer conductor.
  2. While wearing a grounded wrist strap, grasp the outer shell of the cable connector.
  3. Connect the other end of the cable to the test port.
  4. Remove the short from the cable.

Refer to [Chapter 6](#), “Replaceable Parts,” for part numbers and instructions for ordering ESD protection devices.

**Figure 3-1 ESD Protection Setup**



---

## Visual Inspection

Visual inspection and, if necessary, cleaning should be done every time a connection is made. Metal particles from the connector threads may fall into the connector when it is disconnected. One connection made with a dirty or damaged connector can damage both connectors beyond repair.

Magnification is helpful when inspecting connectors, but it is not required and may actually be misleading. Defects and damage that cannot be seen without magnification generally have no effect on electrical or mechanical performance. Magnification is of great use in analyzing the nature and cause of damage and in cleaning connectors, but it is not required for inspection.

### Look for Obvious Defects and Damage First

Examine the connectors first for obvious defects and damage: badly worn plating on the connector interface, deformed threads, or bent, broken, or misaligned center conductors. Connector nuts should move smoothly and be free of burrs, loose metal particles, and rough spots.

### What Causes Connector Wear?

Connector wear is caused by connecting and disconnecting the devices. The more use a connector gets, the faster it wears and degrades. The wear is greatly accelerated when connectors are not kept clean, or are connected incorrectly.

Connector wear eventually degrades performance of the device. Calibration devices should have a long life if their use is on the order of a few times per week. Replace devices with worn connectors.

The test port connectors on the network analyzer test set may have many connections each day, and are therefore more subject to wear. It is recommended that an adapter be used as a test port saver to minimize the wear on the test set's test port connectors.

### Inspect the Mating Plane Surfaces

Flat contact between the connectors at all points on their mating plane surfaces is required for a good connection. See [Figure 2-1 on page 2-3](#). Look especially for deep scratches or dents, and for dirt and metal particles on the connector mating plane surfaces. Also look for signs of damage due to excessive or uneven wear or misalignment.

Light burnishing of the mating plane surfaces is normal, and is evident as light scratches or shallow circular marks distributed more or less uniformly over the mating plane surface. Other small defects and cosmetic imperfections are also normal. None of these affect electrical or mechanical performance.

If a connector shows deep scratches or dents, particles clinging to the mating plane surfaces, or uneven wear, clean and inspect it again. Devices with damaged connectors should be discarded. Determine the cause of damage before connecting a new, undamaged connector in the same configuration.

## Cleaning Connectors

### Inspect Female Connectors

Pay special attention to the contact fingers in the female center conductor. These can be bent or broken, and damage to them is not always easy to see. A connector with damaged contact fingers will not make good electrical contact and must be replaced.

---

**NOTE** This is particularly important when mating nonprecision to precision devices.

---

---

## Cleaning Connectors

Clean connectors are essential for ensuring the integrity of RF and microwave coaxial connections.

### 1. Use Compressed Air or Nitrogen

---

**WARNING** Always use protective eyewear when using compressed air or nitrogen.

---

Use compressed air (or nitrogen) to loosen particles on the connector mating plane surfaces. Clean air cannot damage a connector or leave particles or residues behind.

You can use any source of clean, dry, low-pressure compressed air or nitrogen that has an effective oil-vapor filter and liquid condensation trap placed just before the outlet hose.

Ground the hose nozzle to prevent electrostatic discharge, and set the air pressure to less than 414 kPa (60 psi) to control the velocity of the air stream. High-velocity streams of compressed air can cause electrostatic effects when directed into a connector. These electrostatic effects can damage the device. Refer to **“Electrostatic Discharge”** on page 3-2 for additional information.

### 2. Clean the Connector Threads

---

**WARNING** Keep isopropyl alcohol away from heat, sparks, and flame. Store in a tightly closed container. It is extremely flammable. In case of fire, use alcohol foam, dry chemical, or carbon dioxide; water may be ineffective.

Use isopropyl alcohol with adequate ventilation and avoid contact with eyes, skin, and clothing. It causes skin irritation, may cause eye damage, and is harmful if swallowed or inhaled. It may be harmful if absorbed through the skin. Wash thoroughly after handling.

In case of spill, soak up with sand or earth. Flush spill area with water.

Dispose of isopropyl alcohol in accordance with all applicable federal, state, and local environmental regulations.

---

Use a lint-free swab or cleaning cloth moistened with isopropyl alcohol to remove any dirt or stubborn contaminants on a connector that cannot be removed with compressed air or nitrogen. Refer to [Table 6-2 on page 6-5](#) for part numbers for cleaning swabs.

- a. Apply a small amount of isopropyl alcohol to a lint-free cleaning swab.
- b. Clean the connector threads.
- c. Let the alcohol evaporate, then blow the threads dry with a gentle stream of clean, low-pressure compressed air or nitrogen. Always completely dry a connector before you reassemble or use it.

### 3. Clean the Mating Plane Surfaces

- a. Apply a small amount of isopropyl alcohol to a lint-free cleaning swab.
- b. Clean the center and outer conductor mating plane surfaces. Refer to [Figure 2-1 on page 2-3](#). When cleaning a female connector, avoid snagging the swab on the center conductor contact fingers by using short strokes.
- c. Let the alcohol evaporate, then blow the connector dry with a gentle stream of clean, low-pressure compressed air or nitrogen. Always completely dry a connector before you reassemble or use it.

### 4. Reinspect

Inspect the connector again to make sure that no particles or residue are present.

---

## Connections

Good connections require a skilled operator. *The most common cause of measurement error is bad connections.* The following procedures illustrate how to make good connections.

### How to Make a Connection

#### Preliminary Connection

1. Ground yourself and all devices. Wear a grounded wrist strap and work on a grounded, conductive table mat. Refer to [“Electrostatic Discharge” on page 3-2](#) for ESD precautions.
2. Visually inspect the connectors. Refer to [“Visual Inspection” on page 3-3](#).
3. If necessary, clean the connectors. Refer to [“Cleaning Connectors” on page 3-4](#).
4. Carefully align the connectors. The male connector center pin must slip concentrically into the contact finger of the female connector.
5. Push the connectors straight together.

---

**CAUTION** Do *not* turn the device body. Only turn the connector nut. Damage to the center conductor can occur if the device body is twisted.

---

Do *not* twist or screw the connectors together. As the center conductors mate, there is usually a slight

**Connections**

resistance.

6. The preliminary connection is tight enough when the mating plane surfaces make uniform, light contact. Do not overtighten this connection.

A connection in which the outer conductors make gentle contact at all points on both mating surfaces is sufficient. Very light finger pressure is enough to accomplish this.

7. Make sure the connectors are properly supported. Relieve any side pressure on the connection from long or heavy devices or cables.

**Final Connection Using a Torque Wrench**

1. Use a torque wrench to make a final connection. **Table 3-1** provides information about the torque wrench recommended for use with these calibration kits. A torque wrench is included in the 85038A calibration kit but not in the 85038F/M calibration kits. Refer to **Chapter 6** for part number and ordering information.


**Table 3-1 Torque Wrench Information**

Connector Type	Torque Setting	Torque Tolerance
7-16	226 N-cm (20 in-lb)	$\pm 22.6$ N-cm ( $\pm 2.0$ in-lb)

Using a torque wrench guarantees that the connection is not too tight, preventing possible connector damage. It also guarantees that all connections are equally tight each time.

2. Prevent the rotation of anything other than the connector nut that you are tightening. It may be possible to do this by hand if one of the connectors is fixed (as on a test port). In all situations, however, it is recommended that you use an open-end wrench to keep the body of the device from turning. Refer to **Chapter 6** for part number and ordering information.
3. Position both wrenches within 90 degrees of each other before applying force. See **Figure 3-2**. Wrenches opposing each other (greater than 90 degrees apart) will cause a lifting action which can misalign and stress the connections of the devices involved. This is especially true when several devices are connected together.


**Figure 3-2 Wrench Positions**



wj56f

4. Hold the torque wrench lightly, at the end of the handle only (beyond the groove). See **Figure 3-3**.

**Figure 3-3 Using the Torque Wrench**



wj68d

5. Apply force perpendicular to the wrench handle. This applies torque to the connection through the wrench.  
Do not hold the wrench so tightly that you push the handle straight down along its length rather than pivoting it, otherwise you apply an unknown amount of torque.
6. Tighten the connection just to the torque wrench break point. The wrench handle gives way at its internal pivot point. See **Figure 3-3**. Do not tighten the connection further.

---

**CAUTION** You don't have to fully break the handle of the torque wrench to reach the specified torque; doing so can cause the handle to kick back and loosen the connection. Any give at all in the handle is sufficient torque.

---

## Handling and Storage

Do not pivot the wrench handle on your thumb or other fingers, otherwise you apply an unknown amount of torque to the connection when the wrench reaches its break point.

Do not twist the head of the wrench relative to the outer conductor mating plane. If you do, you apply more than the recommended torque.

## How to Separate a Connection

To avoid lateral (bending) force on the connector mating plane surfaces, always support the devices and connections.

---

**CAUTION** Do *not* turn the device body. Only turn the connector nut. Damage to the center conductor can occur if the device body is twisted.

---

1. Use an open-end wrench to prevent the device body from turning.
2. Use another open-end wrench to loosen the connector nut.
3. Complete the separation by hand, turning only the connector nut.
4. Pull the connectors straight apart without twisting, rocking, or bending either of the connectors.

---

## Handling and Storage

- Install the protective end caps and store the calibration devices in the foam-lined storage case when not in use.
- Never store connectors loose in a box, or in a desk or bench drawer. This is the most common cause of connector damage during storage.
- Keep connectors clean.
- Do not touch mating plane surfaces. Natural skin oils and microscopic particles of dirt are easily transferred to a connector interface and are very difficult to remove.
- Do not set connectors contact-end down on a hard surface. The plating and the mating plane surfaces can be damaged if the interface comes in contact with any hard surface.

## **4 Performance Verification**

## **Introduction**

The performance of your calibration kit can only be verified by returning the kit to Keysight Technologies for recertification. The equipment required to verify the specifications of the devices in the kit has been specially manufactured and is not commercially available.

---

## **How Keysight Verifies the Devices in Your Kit**

Keysight verifies the specifications of these devices as follows:

1. The residual microwave error terms of the test system are verified with precision airlines and shorts that are directly traced to the National Institute of Standards and Technology (NIST). The airline and short characteristics are developed from mechanical measurements. The mechanical measurements and material properties are carefully modeled to give very accurate electrical representation. The mechanical measurements are then traced to NIST through various plug and ring gages and other mechanical measurements.
2. Each calibration device is electrically tested on this system. For the initial (before sale) testing of the calibration devices, Keysight includes the test measurement uncertainty as a guardband to guarantee each device meets the published specification. For recertifications (after sale), no guardband is used and the measured data is compared directly with the specification to determine the pass or fail status. The measurement uncertainty for each device is, however, recorded in the calibration report that accompanies recertified kits.

These two steps establish a traceable link to NIST for Keysight to the extent allowed by the institute's calibration facility. The specifications data provided for the devices in your kit is traceable to NIST through Keysight Technologies.

---

## Recertification

The following will be provided with a recertified kit:

- a new calibration sticker affixed to the case
- a certificate of calibration
- a calibration report for each device in the kit listing measured values, specifications, and uncertainties

---

**NOTE** A list of NIST traceable numbers may be purchased upon request to be included in the calibration report.

---

Keysight Technologies offers a *Standard* calibration for the recertification of your kit. For more information, contact Keysight Technologies. For contact information, see ["Contacting Keysight" on page 5-5](#).

### How Often to Recertify

The suggested initial interval for recertification is 12 months or sooner. The actual need for recertification depends on the use of the kit. After reviewing the results of the initial recertification, you may establish a different recertification interval that reflects the usage and wear of the kit.

---

**NOTE** The recertification interval should begin on the date the kit is *first used* after the recertification date.

---

---

## Where to Send a Kit for Recertification

Contact Keysight Technologies for information on where to send your kit for recertification. For contact information, see ["Contacting Keysight" on page 5-5](#).

When you return the kit, complete and attach a service tag. Refer to ["Returning a Kit or Device to Keysight," on page 5-4](#) for details.


Performance Verification  
**Recertification**

## **5 Troubleshooting**

## Troubleshooting Process

If you suspect a bad calibration, or if your network analyzer does not pass performance verification, follow the steps in [Figure 5-1](#).

Figure 5-1 Troubleshooting Flowchart



pn43m

## **Where to Look for More Information**

This manual contains limited information about network analyzer system operation. For detailed information on using a VNA, ENA or PNA series network analyzer, refer to the appropriate user guide or online Help.

- To view the ENA or PNA online Help, press the Help key on the front panel of the network analyzer.
- To view an online VNA user guide, use the following steps:
  1. Go to [www.keysight.com](http://www.keysight.com).
  2. Enter your VNA model number (Ex: N5242A) in the Search box and click **Search**.
  3. Under the heading **Manuals & Guides**, click on the title/hyperlink for the document PDF you want to view.

If you need additional information, see [“Contacting Keysight” on page 5-5](#).

## **Returning a Kit or Device to Keysight**

If your kit or device requires service, contact Keysight Technologies for information on where to send it. See ["Contacting Keysight," on page 5-5](#). Include a service tag (located near the end of this manual) on which you provide the following information:

- your company name and address
- a technical contact person within your company, and the person's complete phone number
- the model number and serial number of the kit
- the part number and serial number of each device
- the type of service required
- a *detailed* description of the problem and how the device was being used when the problem occurred (such as calibration or measurement)

## **Contacting Keysight**

Assistance with test and measurements needs and information on finding a local Keysight office are available on the Web at:

[www.keysight.com/find/assist](http://www.keysight.com/find/assist)

If you do not have access to the Internet, please contact your Keysight field engineer.

---

**NOTE**      In any correspondence or telephone conversation, refer to the Keysight product by its model number and full serial number. With this information, the Keysight representative can determine whether your product is still within its warranty period.

---



## **6 Replaceable Parts**


## Introduction

**Table 6-1** lists the replacement part numbers for items included in the 85038A/F/M calibration kits and **Figure 6-1** illustrates each of these items.

**Table 6-2 on page 6-5** lists the replacement part numbers for items not included in the calibration kits that are either required or recommended for successful operation of the kits. The 7-16 Adapter Kits can be purchased from Maury Microwave Corp. Get ordering information at [www.maurymw.com](http://www.maurymw.com)

To order a listed part, note the description, the part number, and the quantity desired. Telephone or send your order to Keysight Technologies. For contact information, see “**Contacting Keysight**” on page 5-5.

Figure 6-1 Replaceable Parts for the 85038A/F/M Calibration Kits



**Table 6-1 Replaceable Parts for the 85038A/F/M Calibration Kits**

Item No.	Description	Part of 85038A	Part of 85038F	Part of 85038M	Qty per kit	Keysight Part Number
<b>Calibration Devices (7-16)<sup>a</sup></b>						
1	Female open	✓	✓		1	85038-80002
2	Female short	✓	✓		1	85038-80004
3	Female load	✓	✓		1	85038-80006
4	Male open	✓		✓	1	85038-80003
5	Male short	✓		✓	1	85038-80005
6	Male load	✓		✓	1	85038-80007
<b>Adapters (7-16)<sup>a</sup></b> Purchase from Maury Microwave Corp. Get ordering information at <a href="http://www.maurymw.com">www.maurymw.com</a>						
7	Male to male			✓	1	-
8	Female to female		✓		1	-
<b>Calibration Kit Manual</b>						
9	User's and service guide <sup>b</sup>	✓	✓	✓	1	85038-90010
<b>Wrenches</b>						
10	Open-end wrench	✓			1	8710-2174
11	Torque wrench	✓			1	8710-2175
<b>Protective End Caps for Connectors</b>						
12	For male devices	✓		✓	as required	1401-0418
13	For female devices	✓	✓		as required	1401-0417

a. See "Clarifying the Terminology of a Connector Interface" on page 1-5.

b. Refer to "Printing Copies of Documentation from the Web" on page -iii.

**Table 6-2 Replaceable Parts—Items Not Included in the Calibration Kits**

Description	Qty	Keysight Part Number
<b>Adapter Kits</b> Purchase from Maury Microwave Corp. Get ordering information at <a href="http://www.maurymw.com">www.maurymw.com</a>		
7-16 to 7-16 (includes the following three adapters:)	1	-
7-16-male to 7-16-male	1	-
7-16-female to 7-16-female	1	-
7-16-male to 7-16-female	1	-
7-16 to type-N (includes the following four adapters:)	1	-
7-16-male to type-N-male	1	-
7-16-female to type-N-female	1	-
7-16-male to type-N-female	1	-
7-16-female to type-N-male	1	-
7-16 to 7-mm (includes the following two adapters:)	1	-
7-16-male to 7-mm	1	-
7-16-female to 7-mm	1	-
7-16 to 3.5-mm (includes the following four adapters:)	1	-
7-16-male to 3.5-mm-male	1	-
7-16-female to 3.5-mm-male	1	-
7-16-male to 3.5-mm female	1	-
7-16-female to 3.5-mm female	1	-
<b>ESD Protection Devices</b>		
Grounding wrist strap	1	9300-1367
5 ft grounding cord for wrist strap	1	9300-0980
2 ft by 4 ft conductive table mat with 15 ft grounding wire	1	9300-0797
ESD heel strap	1	9300-1308
<b>Connector Cleaning Supplies</b>		
Anhydrous isopropyl alcohol (>92% pure) <sup>a</sup>	--	--
Foam-tipped cleaning swabs	100	9301-1243

a. Keysight can no longer safely ship isopropyl alcohol, so customers should purchase it locally.

Replaceable Parts  
**Introduction**

# **A Standard Definitions**

## **Class Assignments and Standard Definitions Values are Available on the Web**

Class assignments and standard definitions may change as more accurate model and calibration methods are developed. You can download the most recent class assignments and standard definitions from Keysight's Calibration Kit Definitions Web page at <http://na.support.keysight.com/pna/caldefs/stddefs.html>.

For a detailed discussion of calibration kits, refer to the Keysight Application Note, "Specifying Calibration Standards and Kits for Keysight Vector Network Analyzers." This application note covers calibration standard definitions, calibration kit content and its structure requirements for Keysight vector network analyzers. It also provides some examples of how to set up a new calibration kit and how to modify an existing calibration kit definition file. To download a free copy, go to [www.keysight.com](http://www.keysight.com) and enter literature number 5989-4840EN in the Search window.

## Numerics

**8510 network analyzer**, 1-2

## A

### adapter kits

part numbers, 6-4

### adapters

part numbers, 6-3

### agreements

customer assistance, -ii  
maintenance, -ii

### alcohol

isopropyl  
as cleaning solvent, 3-4  
precautions for use of, 3-4

### altitude

specifications, 2-2

### assistance

customer, -ii

## C

### cal kit

performance  
verifying, 4-2

### calibration

bad, 5-2  
certificate of, 4-3  
constants, 1-2  
permanently stored, 1-2  
frequency, 1-5  
MIL-STD 45662, 4-3  
report, 4-3  
temperature, 2-2  
when to perform, 1-5

### calibration constants

entering, 1-2  
permanently stored, 1-2

### calibration definitions, 1-2

### calibration kit

Keysight Application Note, A-2  
contents, 1-2, 6-2  
drawing of, 6-2  
modifying definition files, A-2  
overview, 1-2  
performance  
how Keysight verifies, 4-2  
verifying, 4-2

### calibration report, 4-3

### certificate of calibration, 4-3

### certification

of device specifications, 2-5

### characteristics

mechanical, 2-3

### class assignments

downloading from Keysight Web site,  
A-2

### cleaning connectors, 3-4

### cleaning supplies, 1-2

part number, 6-4

### cleaning swabs, 3-4

part number, 6-4

### compressed air

for cleaning, 3-4

### connections, 3-2, 3-5

ESD protection, 3-5

final, 3-5

preliminary, 3-5

separating, 3-8

using torque wrench, 3-5

### connector

cleaning, 3-4

damage, 3-3

defects, 3-3

female, 3-3

mating plane surfaces, 3-4

cleaning, 3-4

terminology, 1-5

threads

cleaning, 3-4

inspecting, 3-3

visual inspection, 3-3

wear, 3-3

affect on electrical performance, 3-3

### constants

calibration, 1-2

entering, 1-2

permanently stored, 1-2

### contacting Keysight Technologies, 5-4

### contents

calibration kit, 6-2

drawing of, 6-2

incomplete

what to do, 1-3

## D

### damage

device, 3-3

inspecting for, 3-3

to connectors, 3-3

what to do, 1-3

### damaged connectors, 3-3

### data

recertification, 4-3

### defective connectors, 3-3

### defects

connector, 3-3

### deviation from nominal phase, 2-4

### device

connecting, 3-5

damage, 3-3

disconnecting, 3-8

handling, 3-8

maintenance, 1-5

performance

verifying, 4-2

specifications, 2-4

certification of, 2-5

traceability, 4-2, 4-3

storage, 3-8

temperature, 2-2

visual inspection, 3-3

### disconnections, 3-8

### downloading class assignments & std definitions from the Web, A-2

## E

### electrical specifications, 2-4

### electrostatic discharge, 3-2

supplies

part numbers, 6-4

### end caps

part numbers, 6-3

### environmental

regulations, 3-4

requirements, 2-2

specifications, 2-2

### equipment required, 1-2

### ESD, 3-2

precautions, 3-2, 3-4

supplies

part numbers, 6-4

## F

### female connectors, 3-3

inspection of, 3-3

### frequency

specifications, 2-4

### frequency of calibration, 1-5

## G

### gage

connector, 1-2

## H

### handling, 3-8

### how often to calibrate, 1-5

### humidity

specifications, 2-2

## I

**incoming inspection**, 1-3  
**information, troubleshooting**, 5-3  
**inspection**  
  damage, 3-3  
  defects, 3-3  
  female connectors, 3-3  
  incoming, 1-3  
  mating plane surfaces, 3-3  
  visual, 3-3  
**isopropyl alcohol**  
  as cleaning solvent, 3-4  
  precautions for use of, 3-4

## K

**Keysight Technologies**  
  application note, A-2  
  contacting, 5-4, 5-5  
**kit**  
  contents, 1-2, 6-2  
  drawing of, 6-2  
  overview, 1-2

## L

**loads**  
  part numbers, 6-3  
  specifications, 2-4

## M

**maintenance**, 3-2  
  agreements, -ii  
  preventive, 1-5  
**maintenance of devices**, 1-5  
**making connections**, 3-5  
  ESD protection, 3-5  
  precautions, 3-5  
**manual**  
  printing, -iii  
**mating plane surfaces**  
  cleaning, 3-4  
  connector, 3-4  
  inspection of, 3-3  
**mechanical characteristics**, 2-3  
  affect on electrical performance, 2-3  
**MIL-STD 45662**  
  calibration, 4-3  
**modifying calibration kit definition files**, A-2

## N

**National Institute of Standards and Technology (NIST)**, 2-5, 4-2  
**network analyzer, 8510**, 1-2  
**NIST (National Institute of Standards and Technology)**, 2-5, 4-2  
**nitrogen**  
  for cleaning, 3-4  
**numbers**  
  serial, 1-4

## O

**open-end wrench**, 3-8  
  part number, 6-3  
**opens**  
  part numbers, 6-3  
  specifications, 2-4  
**ordering**  
  parts, 6-2

## P

**part numbers**, 6-2  
  of items in kit, 6-3  
  of items not in kit, 6-4  
**parts**  
  not included in kit, 6-2  
  ordering, 6-2  
  replaceable, 6-2  
**performance verification**  
  fail, 5-2  
**permanently stored calibration definitions**, 1-2  
**pin depth**, 2-3  
  protrusion, 2-3  
  recession, 2-3  
**preventive maintenance**, 1-5  
**printing the manual**, -iii  
**protective end caps**  
  part numbers, 6-3  
**protrusion**  
  pin depth, 2-3

## R

**recertification**  
  how to order, 4-3  
  interval, 4-3  
  what's included, 4-3  
  where it's done, 4-3  
**recession**  
  pin depth, 2-3  
**regulations**  
  environmental, 3-4

**replaceable parts**, 6-2  
  drawing of, 6-2  
**requirements**  
  environmental, 2-2  
**return**  
  kit or device to Keysight, 5-4  
**return loss**  
  specifications, 2-4

## S

**separating connections**, 3-8  
**serial numbers**, 1-4  
  devices, 1-4  
  recording, 1-4  
**service**, 5-4  
  tag, 1-3, 4-3  
**service tag**, 5-4  
**shorts**  
  part numbers, 6-3  
  specifications, 2-4  
**specifications**, 2-2  
  altitude  
    operating, 2-2  
    storage, 2-2  
  certification of, 2-5  
  deviation from nominal phase, 2-4  
  device, 2-4  
  electrical, 2-4  
  environmental, 2-2  
  frequency, 2-4  
  humidity, 2-2  
  return loss, 2-4  
  temperature, 2-2  
  torque wrench, 3-5  
  traceability, 4-2, 4-3  
**standard definitions**  
  downloading from Keysight Web site, A-2  
**standards**  
  international, 2-5, 4-2  
  National Institute of Standards and Technology (NIST), 2-5, 4-2  
**static discharge**, 3-2  
**storage**, 3-8  
  altitude, 2-2  
  temperature, 2-2  
**supplies**  
  cleaning, 1-2  
  part number, 6-4  
**swabs**  
  cleaning, 3-4

## T

**tag**

---

# Index

---

service, 1-3, 4-3, 5-4

## **temperature**

affect on electrical performance, 2-2

calibration, 2-2

cautions about, 2-2

changes in, 2-2

device, 2-2

error-corrected, 2-2

measurement, 2-2

specifications, 2-2

operating, 2-2

storage, 2-2

verification and measurement, 2-2

## **terminology, connector, 1-5**

## **test data, 4-3**

## **threads**

connector

cleaning, 3-4

inspecting, 3-4

## **torque wrench, 1-2**

part number, 6-3

specifications, 3-5

## **traceability**

of device specifications, 4-2, 4-3

## **troubleshooting, 5-2**

# U

## **user's and service guide**

part number, 6-3

# V

## **verification**

temperature, 2-2

## **visual inspection, 3-3**

# W

## **wear**

connector, 3-3

affect on electrical performance, 3-3

## **when to calibrate, 1-5**

## **wrench**

open-end, 1-2, 3-6, 3-8

part number, 6-3

proper positioning of, 3-6

torque, 1-2, 3-5, 3-6, 3-7

part number, 6-3

precautions for use of, 3-7

proper use of, 3-7



This information is subject to change without notice.  
© Keysight Technologies 1999 - 2015  
February 2015



85038-90010  
[www.keysight.com](http://www.keysight.com)