

# Keysight 87421/2A Power Supply

# Notices

## Copyright Notice

© Keysight Technologies 1992 - 2019

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies as governed by United States and international copyright laws.

## Manual Part Number

87421-90001

## Edition

Edition 5, March 19, 2019

## Printed in:

Printed in Malaysia

## Published by:

Keysight Technologies  
Bayan Lepas Free Industrial Zone,  
11900 Penang, Malaysia

## Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

## Declaration of Conformity

Declarations of Conformity for this product and for other Keysight products may be downloaded from the Web. Go to <http://www.keysight.com/go/conformity>. You can then search by product number to find the latest Declaration of Conformity.

## U.S. Government Rights

The Software is "commercial computer software," as defined by Federal Acquisition Regulation ("FAR") 2.101. Pursuant to FAR 12.212 and 27.405-3 and Department of Defense FAR Supplement ("DFARS") 227.7202, the U.S. government acquires commercial computer software under the same terms by which the software is customarily provided to the public. Accordingly, Keysight provides the Software to U.S. government customers under its standard commercial license, which is embodied in its End User License Agreement (EULA), a copy of which can be found at <http://www.keysight.com/find/sweula>. The license set forth in the EULA represents the exclusive authority by which the U.S. government may use, modify, distribute, or disclose the Software. The EULA and the license set forth therein, does not require or permit, among other things, that Keysight: (1) Furnish technical information related to commercial computer software or commercial computer software documentation that is not customarily provided to the public; or (2) Relinquish to, or otherwise provide, the government customers in excess of these rights customarily provided to the public to use, modify, reproduce, release, perform, display, or disclose commercial computer software or commercial computer software documentation. No additional government requirements beyond those set forth in the EULA shall apply, except to the extent that those terms, rights, or licenses are explicitly required from all providers of commercial computer software pursuant to the FAR and the DFARS and are set forth specifically in writing elsewhere in the EULA. Keysight shall be under no obligation to update, revise or otherwise modify the Software. With respect to any technical data as defined by FAR 2.101, pursuant to FAR 12.211 and 27.404.2 and DFARS 227.7102, the U.S. government acquires no greater than Limited Rights as defined in FAR 27.401 or DFAR 227.7103-5 (c), as applicable in any technical data.

## Warranty

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, KEYSIGHT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED, WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. KEYSIGHT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR OF ANY INFORMATION CONTAINED HEREIN. SHOULD KEYSIGHT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS DOCUMENT THAT CONFLICT WITH THESE TERMS, THE WARRANTY TERMS IN THE SEPARATE AGREEMENT SHALL CONTROL.

## Safety Information

### CAUTION

A CAUTION notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

### WARNING

A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.







## Printing Copies of Documentation from the Web

To print copies of documentation from the Web, download the PDF file from the Keysight web site:

- Go to <http://www.keysight.com>.
- Enter the document's part number (located on the front page) in the **Quick Search** box.
- Click **GO**.
- Click on the hyperlink for the document.
- Click the printer icon located in the tool bar.

## Safety Symbols

The following symbols on the instrument and in the documentation indicate precautions which must be taken to maintain safe operation of the instrument.

	When you see this symbol on your instrument, you should refer to the instrument's instruction manual for important information.		This symbol indicates hazardous voltages.
	The laser radiation symbol is marked on products that have a laser output.		This symbol indicates that the instrument requires alternating current (ac) input.
	This symbol indicates that the power line switch is ON.		This symbol indicates that the power line switch is OFF or in STANDBY position.

## Safety Considerations

Read the information below before using this instrument.

The following general safety precautions must be observed during all phases of operation, service, and repair of this instrument. Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards for design, manufacture, and intended use of the instrument. Keysight Technologies assumes no liability for the customer's failure to comply with these requirements.

### WARNING

#### SAFETY EARTH GROUND



This is a Safety Class I product (provided with a protective earthing terminal). An uninterruptible safety earth ground must be provided from the main power source to the product input wiring terminals, power cord, or supplied power cord set. Whenever it is likely that the protection has been impaired, the product must be made inoperative and secured against any unintended operation.

### CAUTION

#### BEFORE APPLYING POWER

Verify that the product is configured to match the available main power source as described in the input power configuration instructions in this manual. If this product is to be powered by autotransformer, make sure the common terminal is connected to the neutral (grounded) side of the ac power supply.

## Regulatory Markings

	The CE mark is a registered trademark of the European Community. This CE mark shows that the product complies with all the relevant European Legal Directives.		The CSA mark is a registered trademark of the Canadian Standards Association.
<b>ISM 1-A</b>	This text indicates that the instrument is an Industrial Scientific and Medical Group 1 Class A product (CISPER 11, Clause 4).		

## Waste Electrical and Electronic Equipment (WEEE) Directive

This instrument complies with the WEEE Directive marking requirement. This affixed product label indicates that you must not discard this electrical or electronic product in domestic household waste.

### Product category:

With reference to the equipment types in the WEEE directive Annex 1, this instrument is classified as a “Monitoring and Control Instrument” product.

The affixed product label is as shown below.



Do not dispose in domestic household waste.

To return this unwanted instrument, contact your nearest Keysight Service Center, or visit <http://about.keysight.com/en/companyinfo/environment/takeback.shtml> for more information.

## Sales and Technical Support

To contact Keysight for sales and technical support, refer to the support links on the following Keysight websites:

- [www.keysight.com/find/amplifiers](http://www.keysight.com/find/amplifiers)  
(product-specific information and support, software and documentation updates)
- [www.keysight.com/find/assist](http://www.keysight.com/find/assist)  
(worldwide contact information for repair and service)

THIS PAGE HAS BEEN INTENTIONALLY LEFT BLANK.

# Table of Contents

Printing Copies of Documentation from the Web	3
Safety Symbols	3
Safety Considerations	4
Regulatory Markings	4
Waste Electrical and Electronic Equipment (WEEE) Directive	5
Product category:	5
Sales and Technical Support	5
List of Figures	9
List of Tables	11
<b>1 Introduction</b>	
Product Overview	14
87421A power supply	14
87422A power supply	14
<b>2 Specifications</b>	
Environmental Specifications	16
<b>3 Operating and Service</b>	
Operating and Service	18
Operating instructions	18
Fuse replacement	18
Service information	18

THIS PAGE HAS BEEN INTENTIONALLY LEFT BLANK.



List of Figures

Figure 1-1      87421A and 87422A power supply ..... 14

THIS PAGE HAS BEEN INTENTIONALLY LEFT BLANK.

List of Tables

Table 2-1	Environmental specifications	16
-----------	------------------------------	----

THIS PAGE HAS BEEN INTENTIONALLY LEFT BLANK.

# 1 Introduction

Product Overview 14

This chapter provides you the overview of Keysight 87421A and 87422A power supply.

## Product Overview

The 87421A and 87422A power supplies are universal input, dc switching power supplies designed to provide the bias voltages needed to operate microwave system amplifiers.



**Figure 1-1** 87421A and 87422A power supply

### 87421A power supply

The 87421A is a 25watt,  $\pm 12\text{Vdc}$  power supply designed to power the 83006A, 83017A, and 87415A microwave system amplifiers.

- The power supply includes an AC power cord with protective ground for connection to 100–240Vac.
- Also included is a 2 meter cable with a 3-pin connector on one end and a D-subminiature connector (83006–60005) on the other end for direct connection to the microwave system amplifiers.

The 87421A may also be used to provide bias voltages for the 83440C high speed lightwave converter. Using the 83440–60009 lightwave converter power supply adapter, the 87421A becomes a compact convenient power supply for use with the lightwave converter.

### 87422A power supply

The 87422A is a 90watt,  $\pm 15\text{Vdc}$  power supply designed to power the 83020A microwave system power amplifier. One additional  $\pm 12\text{V}$  dc output is provided to power a preamplifier such as the 83006A or 83017A.

The power supply includes an AC power cord with protective ground for connection to 100–240Vac and two 2 meter cables:

- one cable (83006–60005) with a 3-pin connector on one end and a D-subminiature connector on the other end
- one cable (87422–60001) for direct connection to the 83020A amplifier

## 2 Specifications

### Environmental Specifications 16

This chapter provides you the environmental specifications of Keysight 87421A and 87422A power supply.

## Environmental Specifications

**Table 2-1** Environmental specifications

Parameter	Specification
<b>Temperature</b>	
– Operating	0°C to 55°C
– Storage	-40°C to 70°C
<b>Humidity</b>	
– Operating	15% to 95% RH at 40°C
– Non operating supersoak	90% RH at 65°C for 24 hours
– Condensation	After stabilizing at 50% RH/25°C, product shall be subjected to 95% RH at 40°C. Product shall suffer no damage and operate within specifications within 15min.
<b>Shock</b>	
– End use handling shock	Half-sine shock, < 3ms duration, 155in/s $\Delta V$ , minimum 3 shocks each face
– Transportation shock (Type 1)	Trapezoid shock, 30g acceleration, 292in/s $\Delta V$ , minimum 3 shocks each face
<b>Vibration</b>	
– Operating	Random: 0.0001 $g^2/Hz$ for 5Hz to 350Hz, -6dB/octave for 350Hz to 500Hz, minimum 10min/axis for all 3 axes
– Survival	Random: 0.015 $g^2/Hz$ from 5Hz to 100Hz, -6dB/octave from 100Hz to 137Hz, 0.008 $g^2/Hz$ from 137Hz to 350Hz, -6dB/octave from 350Hz to 500Hz, minimum 10min/axis for all 3 axes Swept Sine: Resonant search from 5Hz to 500Hz, 1 octave/min sweep rate, 5 min resonant dwell at 4 resonances/axis
<b>Altitude</b>	
– Operating	4,600m
– Storage	15,300m
<b>ESD immunity</b>	
– Direct discharge	4kV
– Air discharge	8kV



## 3 Operating and Service

Operating and Service 18

This chapter provides you the operating instructions and service information.

## Operating and Service

### Operating instructions

To operate the power supplies, connect the power supplies output to the amplifier dc input with the cable provided. The power supplies include a power cord with protective ground for connection to 100-240Vac.

Mounting holes are provided. The power supplies may be mounted in any orientation with user supplied hardware.

#### NOTE

- To ensure continued compliance to product safety and EMC specifications, connect the power supplies to an amplifier only with the cable provided.
  - Do not attempt to modify or extend the dc output cable. Contact your local Keysight Technologies Sales and Service office for help with special applications requiring longer cables.
- 

### Fuse replacement

The power supplies have a fuse located next to the IEC appliance inlet. If it is necessary to replace the fuse, disconnect the power supply from both AC power and the amplifier. Using a slotted screw driver, open the fuse cover and replace the fuse.

### Service information

The power supplies contain no user serviceable parts (with the exception of the fuse). If service is required, contact your nearest Keysight Technologies Service Center. Refer to [“Sales and Technical Support”](#) on page 5.

Available instruments and accessories

- 83006A microwave system amplifier 0.01 - 26.5GHz
- 83017A microwave system amplifier 0.5 - 26.5GHz
- 83018A microwave system amplifier 2 - 26.5GHz
- 83020A microwave system power amplifier 2 - 26.5GHz
- 87415A microwave system amplifier 2 - 8GHz
- 87405C microwave system pre-amplifier 0.1 - 18GHz
- 83440-60009 lightwave converter power supply adapter



This information is subject to change without notice. Always refer to the Keysight website for the latest revision.

© Keysight Technologies 1992 - 2019  
Edition 5, March 19, 2019

Printed in Malaysia



87421-90001

[www.keysight.com](http://www.keysight.com)